

# IMPACT IMPACT IMPACT Test Equipment

## Blaine Air Permeability Apparatus



*to EN 196-6, COMPARABLE TO, ASTM C204,  
AASHTO T153, BS 4359:2, UNI 7374, NF P15-  
442 -UNE 80-106*

Code	CE090
British Standard	BS4359
Other Standards	EN196-6
Approx. Weight (kg)	12
Tariff	90248090

Used to determine the fineness of Portland cement in terms of the specific surface expressed as total surface area in square centimetres per gram of cement.

The apparatus is supplied with a glass U-tube manometer with valve, steel stand, test cell with disk and plunger all in stainless steel, rubber aspirator bulb, 1000 filter paper disks, manometer liquid and accessories.

Weight: 12kg

A user guide is available to download from the Impact web site [www.impact-test.co.uk](http://www.impact-test.co.uk)

Impact Test Equipment Ltd  
Stevenston Industrial Estate  
Stevenston  
Ayrshire  
KA20 3LR  
United Kingdom

Tel: +44 (0) 1294 602626  
Fax: +44 (0) 1294 461168  
Email: [sales@impact-test.co.uk](mailto:sales@impact-test.co.uk)

[www.impact-test.co.uk](http://www.impact-test.co.uk)  
[www.impact-test.com](http://www.impact-test.com)  
[www.impact-testsets.co.uk](http://www.impact-testsets.co.uk)

